

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ0300.01
---	---	-----------

**Technoview® Rivaroxaban CAL Set High 5x1 mL**

REF      5090171  
LOT      0381B00.01  
      2020-06-30

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
1x1 mL	Technoview Rivaroxaban CAL1 High	5181B00	2020-06-30
1x1 mL	Technoview Rivaroxaban CAL2 High	5281B00	2020-06-30
1x1 mL	Technoview Rivaroxaban CAL3 High	5381B00	2020-06-30
1x1 mL	Technoview Rivaroxaban CAL4 High	5481B00	2020-06-30
1x1 mL	Technoview Rivaroxaban CAL5 High	5581B00	2020-06-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	Rivaroxaban Control Medium <span style="border: 1px solid black; padding: 2px;">LOT</span> 3551B01	114.9 – 191.5 ng/mL	142.6	OK
	Rivaroxaban Control High <span style="border: 1px solid black; padding: 2px;">LOT</span> 3381B00	245.2 – 408.6 ng/mL	305.0	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von  
Prepared by**

  
Christoph ENNE

27. JULI 2018  
Datum / Date:

**Freigabe von  
Approved by**

  
Eva Maria HAGER

30. JULI 2018  
Datum / Date: