



# ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQ450D.02

Att.03

4583B0D.01

## F V Mangelplasma immunads, 5x1mL

REF	5134004
LOT	4583B0D.01
	2020-05-31

Analytische Verfahren Analytical Procedures	Anforderung Requirement	Ergebnis Result
visuelle Kontrolle visual inspection	Cap color	blue confirmed OK
	Vial type	clear glass vial confirmed OK
	Vial label	according to packaging SOP confirmed OK
	Vial barcode	according to packaging SOP confirmed OK
	Box label	according to packaging SOP confirmed OK
	Box barcode	according to packaging SOP confirmed OK
	Package insert	present confirmed OK
	Appearance before reconstitution	buff colored plug of lyophilized material confirmed OK
	Appearance after reconstitution	buff colored liquid without visible particles confirmed OK
Virologie / Virology  The testing methods applied were FDA-approved or CE marked.	HBsAg / HBsAg HIV-Ak / HIV-Ab HCV-AK / HCV-Ab  Negative Negative Negative	Negative Negative Negative  OK
Kontrollwiederfindung FV am Ceveron® Alpha  Recovery of controls FV on Ceveron® Alpha	<b>Control N</b> LOT 1P71B01  <b>Control A</b> LOT 3P71B01	82.5 - 137.5 % 130.2 %  18.8 - 31.3 % 27.8 %  OK OK

Durchgeführt von  
Prepared by

**Christoph ENNE**

17. JULI 2018  
Datum / Date:

Freigabe von  
Approved by

**Eva Maria HAGER**

17. JULI 2018  
Datum / Date: