

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1L00.01 LOT 1L81B00.01
---	---	------------------------------------

Platelet Poor Plasma 5x2mL

REF 5343022
LOT 1L81B00.01
 2020-01-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	7.4 – 7.8	7.7	OK
Restfeuchte Residual moisture	≤1%	0.38	OK
OD _{320nm}	≤10	5.08	OK
Verwendbarkeit Usability Ceveron® alpha	LA Ratio: Technoclot® LA Screen, LOT 6Y61B01 Technoclot® LA Confirm, LOT 6X71B00	0.89 – 1.49	1.12 OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK	

**Durchführung von
Prepared by**


 Lavdrin NASUFI

14. MRZ. 2018
 Datum / Date:

**Freigabe von
Approved by**


 Stefan STEINER

14. MRZ. 2018
 Datum / Date: