



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ2V00.02

Lupus Inhibitorplasma 5x1 mL

REF 5343019

LOT 2V83B00.01

 2020-06-30

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result		
pH	7.4 – 7.8	7.7	OK	
Restfeuchte Residual moisture	≤ 1 %	0.07	OK	
OD _{320nm}	≤ 10	5.2	OK	
Verwendbarkeit Usability Ceveron® alpha	LA Ratio: Technoclot® LA Screen, LOT 6Y61B01 Technoclot® LA Confirm, LOT 6X71B01	1.64 – 2.73	2.31	OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK		

Durchführung von
Prepared by

Suzana GRESAK PODAKAR

20. JULI 2018

Datum / Date:

Freigabe von
Approved by

Christoph ENNE

20. JULI 2018

Datum / Date: