





**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

DQRB00AZ

TECHNOZYM® vWF:CBA ELISA Kit 96T

REF 5450301
LOT RB87B00.02
 2019-11-30

Inhalt Content	Reagenz Reagent	LOT	
1 x 12	vWF:CBA Plate Collagen Typ III	RNC83B0	2020-05-31
1 x 0.3mL	CBA Conjugate	RNE72B0	2019-11-30
1x 80mL	Wash buffer Concentrate	RNA82B0	2020-06-30
1 x 90mL	Incubation buffer	RNB81B0	2020-02-29
1 x 12mL	TMB Substrate	RKN82B0	2020-05-31
1 x 12mL	Stop solution	RKK81B0	2020-04-30
1 x 0.5mL	vWF Calibrator 1	1ND71B0	2019-12-31
1 x 0.5mL	vWF Calibrator 2	2ND71B0	2019-12-31
1 x 0.5mL	vWF Calibrator 3	3ND71B0	2019-12-31
1 x 0.5mL	vWF Calibrator 4	4ND71B0	2019-12-31
1 x 0.5mL	vWF Calibrator 5	5ND71B0	2019-12-31
1 x 0.5mL	vWF High Control	AND71B0	2019-12-31
1 x 0.5mL	vWF Low Control	BND71B0	2019-12-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.49	OK
	Calibrator 5	< 0.1	0.00	OK
Verwendbarkeit Usability	High Control (AND71B0)	0.84 – 1.40 IU/mL	1.11	OK
	Low Control (BND71B0)	0.18 – 0.42 IU/mL	0.29	OK

**Durchführung von
Prepared by**

Fikret NASUFI *Fikret Nasufi*

13. SEP. 2018

Datum / Date:

**Freigabe von
Approved by**

Suzana GRESAK PODAKAR *Suzana Gresak Podakar*

14. SEP. 2018

Datum / Date: