



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRP00AZ

TECHNOZYM® ADAMTS-13 INH ELISA Kit 96T

REF 5450401
LOT RP8AB00.01
2019-08-31

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x 12	Adamts13 INH Plate	RPS88B0	2020-07-31
1 x 0.3mL	Adamts13 INH Conjugate	RPV81B0	2020-04-30
1x 80mL	Wash buffer Concentrate	RNA82B0	2020-06-30
1 x 90mL	Incubation buffer	RNB82B0	2020-06-30
1 x 12mL	TMB Substrate	RKN82B0	2020-05-31
1 x 12mL	Stop solution	RKK81B0	2020-04-30
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH High Control	RPY71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Low Control	RPZ71B0	2019-08-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.75	OK
	Calibrator 5	< 0.1	0.00	OK
Verwendbarkeit Usability	High Control (RPY71B0)	60.24 – 100.40 U/mL	81.99	OK
	Low Control (RPZ71B0)	16.64 – 38.84 U/mL	28.03	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

04. SEP. 2018
Datum / Date:

Freigabe von
Approved by

Michael KAFKA *[Signature]*

04. SEP. 2018
Datum / Date: