

CERTIFICATE OF ANALYSIS

CHROMOGENIX

Instrumentation
Laboratory
A Werfen Company

PRODUCT NAME:	COATEST APC Resistance Control Plasma Level 1	EXPIRATION DATE: 2020-04
PRODUCT NUMBER:	82265063	
LOT	N0386245	

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION	
	MIN	MAX	UNITS			
Analysis with COATEST APC R:						
APT Time	<u>CaCl₂</u>	28	34	seconds	32	PASS
	<u>APC/CaCl₂</u>	90	--	seconds	117	PASS
APC Ratio		3.5	4.4	--	3.7	PASS

Analysis with COATEST APC R V:						
APT Time	<u>CaCl₂</u>	--	38	seconds	36	PASS
	<u>APC/CaCl₂</u>	75	--	seconds	112	PASS
APC Ratio		2.7	3.2	--	3.1	PASS

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: D.Tencza
PRINT

D. Tencza
SIGNATURE

DATE: 4/24/2018 *4/24/18*

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: N.Patel
PRINT

N. Patel
SIGNATURE

DATE: 4/24/2018 *4-24-18*

TITLE: Senior QC Analyst

QCM00C00423L-1, Rev. 01