

CERTIFICATE OF ANALYSIS

CHROMOGENIX


**Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME:	COATEST APC Resistance Control Plasma Level 2	EXPIRATION DATE: 2020-01
PRODUCT NUMBER:	82266863	
LOT NUMBER:	N0185371	

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION	
	MIN	MAX	UNITS			
Analysis with COATEST APC R:						
APT Time	CaCl ₂	27	44	seconds	42	PASS
	APC/CaCl ₂	--	101	seconds	90	PASS
APC Ratio		--	2.3	--	2.1	PASS

Analysis with COATEST APC R V:						
APT Time	CaCl ₂	27	40	seconds	38	PASS
APC Ratio		1.4	1.9	--	1.8	PASS

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: Nehal Patel *N. M. Patel*
PRINT SIGNATURE

DATE: 7/27/2018 *7-27-18*

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: Donna Tencza *Donna Tencza*
PRINT SIGNATURE

DATE: 7/27/2018 *7/27/18*

TITLE: Senior QC Analyst