

CERTIFICATE OF ANALYSIS

CHROMOGENIX

Instrumentation
Laboratory
A Werfen Company

PRODUCT NAME :	COAMATIC FACTOR VIII	EXPIRATION DATE:	2020-06
PRODUCT NUMBER :	82258563		
LOT NUMBER:	N0789231		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	PASS/FAIL
	MIN	MAX	UNITS		
HIGH RANGE					
Slope	0.33	0.57	--	0.43	PASS
Intercept	---	0.12	--	0.05	PASS
r	0.990	---	--	0.999	PASS
LOW RANGE					
Slope	2.8	6.9	--	5.3	PASS
Intercept	---	0.21	--	0.12	PASS
r	0.990	---	--	0.999	PASS

We hereby certify that the above product(s):
 · have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
 · meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: Nehal Patel *N.M. Patel*
 PRINT SIGNATURE

DATE: 8-2-18
 8/2/2018

TITLE: Senior QC Analyst
 APPROVAL SIGNATURE: Donna Tencza *Donna Tencza*
 PRINT SIGNATURE

DATE: 8/2/18
 8/2/2018

TITLE: Senior QC Analyst