
	<p style="text-align: center;">ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</p>	<p style="text-align: center;">DQ7D00AZ</p>
---	---	---

## TECHNOTHROMBIN<sup>®</sup> TGA RC Low 5x0.5mL

REF            5006212

LOT            7D83B00.01

            2020-01-31

Inhalt Content	Reagenz Reagent	LOT	
5x0,5 mL	TGA RC Low	7D83B00	2020-01-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z82B00 254 - 423	340.6	OK
	TGA Control Low (nM Peak Thrombin) LOT 2883B00 90 - 150	141.9	OK

**Durchführung von**  
**Prepared by**

Mikrot NASUFI *Mikrot Nasufi*

25. SEP. 2018

Datum / Date:

**Freigabe von**  
**Approved by**

Michael KAFKA *[Signature]*

26. SEP. 2018

Datum / Date: