

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7G00AZ
---	---	----------

TECHNOTHROMBIN[®] TGA RD 50x1.5mL

REF 5006222

LOT 7G81B00.01

 2020-01-31

Inhalt Content	Reagenz Reagent	LOT	
50x1.5 mL	TGA RD	7G81B00	2020-01-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z82B00	435 - 725	606.2	OK
	TGA Control Low (nM Peak Thrombin) LOT 2883B00	263 - 438	393.0	OK

Durchführung von
Prepared by

Fikrot NASUFI *Fikrot Nasufi*

26. SEP. 2018

Datum / Date:

Freigabe von
Approved by

Michael KAFKA *[Signature]*

26. SEP. 2018

Datum / Date: