
	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ8F00AZ
---	---	----------

## TECHNOTHROMBIN® TGA SUB 50x1.5mL

REF      5006230

LOT      8F71B01.01

      2020-08-31

Inhalt Content	Reagenz Reagent	LOT	
50x1.5mL	TGA Substrat	8F71B01	2020-08-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z82B00	TGA RCL Reagent LOT 7D83B00	254 - 423	338.6	OK
		TGA RCH Reagent LOT 7F81B00	377 - 628	518.2	OK
	TGA Control Low (nM Peak Thrombin) LOT 2883B00	TGA RCL Reagent LOT 7D83B00	90 - 150	120.2	OK
		TGA RCH Reagent LOT 7F81B00	206 - 344	276.3	OK

**Durchführung von**  
**Prepared by**

Fikrot NASUP *Fikrot Nasch*

05. OKT. 2018

Datum / Date:

**Freigabe von**  
**Approved by**

Michael KAFKA

05. OKT. 2018

Datum / Date: