
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1Z00AZ
---	---	----------

TECHNOTHROMBIN[®] TGA Control High 5x1mL

REF 5006320
LOT 1Z82B00.01
 2019-05-31

Inhalt Content	Reagenz Reagent	LOT	
5x1mL	TGA Control High	1Z82B00	2019-05-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Reagenz RC Low (nM Peak Thrombin) LOT 7D83B00	254 - 423	352.9 OK
	TGA Reagenz RC High (nM Peak Thrombin) LOT 7F81B00	377 - 628	576.0 OK

Durchführung von
Prepared by

Elkrot NASUFI *Elkrot Nasufi*

25. SEP. 2018

Datum / Date:

Freigabe von
Approved by

Michael KAFKA *[Signature]*

26. SEP. 2018

Datum / Date: