



| | | |
|---|---|-----------|
|  | ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS | DQ6L0A.01 |
|---|---|-----------|

Siron LIS (aPTT liquid) 10x4 mL

REF 5035119
LOT 6L83BA0.01
 2020-08-31

| Analytische Prozedur Analytical Procedure | Anforderung Requirement | Ergebnis Result | |
|--|----------------------------------|--------------------|------------|
| pH | 7.0 – 7.6 | 7.2 | OK |
| Verwendbarkeit Usability Ceveron® alpha | CCN LOT 1P71B01 | 26.0 – 32.0 sec | 29.8 OK |
| | CCA LOT 3P71B01 | 65.1 – 80.3 sec | 75.9 OK |
| Etiketten und Barcode Überprüfung Label and bar code Verification | | OK | |

**Durchführung von
Prepared by**




Lavdrija NASUFI

0 5. OKT. 2018

Datum / Date:

**Freigabe von
Approved by**



Christoph ENNE

0 5. OKT. 2018

Datum / Date: