
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3300.01
---	---	-----------

TECHNOVIEW® Rivaroxaban Control High 5x1mL

REF 5090174
LOT 3381B00.01
 2020-04-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negativ Negativ Negativ	Negativ Negativ Negativ	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		245.2 – 408.6 ng/mL	326.9	
Verwendbarkeit Usability Ceveron® alpha	TECHNOCHROM® anti-Xa	245.2 – 408.6 ng/mL	332.7	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Christoph ENNE

13. JULI 2018
Datum / Date:

**Freigabe von
Approved by**


Eva Maria HAGER

16. JULI 2018
Datum / Date: