
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0F00.03
---	---	-----------

F VIII Inhibitor Reagent Kit 2-4T.

REF 5152005
LOT 0F85B00.01
 2019-08-31

Inhalt Content	Reagenz Reagent	LOT	
2 x 2.63 mL	FVIII Normal Plasma	1G81B00	2020-06-30
1 x 1 mL	FVIII Inhibitor Plasma	2K71B00	2019-08-31
1 x 1 mL	FVIII Inhibitor Free Plasma	2M81B0D	2020-03-31
1 x 17 mL	Imidazol Puffer	9J74BA0	2022-07-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	FVIII Inhibitor Plasma LOT 2K71B00	11.9 – 19.9 BE/mL	15.2	OK
FVIII clotting, Siron LS Ceveron® alpha	FVIII Inhibitor Free Plasma LOT 2M81B0D	< 0.4 BE/mL	< 0.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Eva Maria HAGER

31. JULI 2018

Datum / Date:

Freigabe von
Approved by


Christoph ENNE

31. JULI 2018

Datum / Date: