
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ2V00.02
---	---	-----------

Lupus Inhibitorplasma Low 5x1 mL

REF 5343021
LOT 2U71B01.01
 2020-05-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	7.4 – 7.8	7.7	OK
Restfeuchte Residual moisture	≤1%	0.26	OK
OD _{320nm}	≤10	6.2	OK
Verwendbarkeit Usability Ceveron® alpha	LA Ratio: Technoclot® LA Screen, LOT 6Y71B00 Technoclot® LA Confirm, LOT 6X71B00	1.21 – 2.01	1.52 OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK	

**Durchführung von
Prepared by**


Stefan STEINER

04. OKT. 2018
Datum / Date:

**Freigabe von
Approved by**


Eya Maria HAGER

04. OKT. 2018
Datum / Date: