
	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQRB00AZ
---	---	----------

**TECHNOZYM® vWF:CBA ELISA Kit 96T**

**REF**      5450301  
**LOT**      RB89B00.01  
      2019-11-30

Inhalt Content	Reagenz Reagent	<b>LOT</b>	
1 x 12	vWF:CBA Plate Collagen Typ III	RNC84B0	2020-08-31
1 x 0.3mL	CBA Conjugate	RNE72B0	2019-11-30
1x 80mL	Wash buffer Concentrate	RNA82B0	2020-06-30
1 x 90mL	Incubation buffer	RNB82B0	2020-06-30
1 x 12mL	TMB Substrate	RKN82B0	2020-05-31
1 x 12mL	Stop solution	RKK82B0	2020-08-31
1 x 0.5mL	vWF Calibrator 1	1ND71B0	2019-12-31
1 x 0.5mL	vWF Calibrator 2	2ND71B0	2019-12-31
1 x 0.5mL	vWF Calibrator 3	3ND71B0	2019-12-31
1 x 0.5mL	vWF Calibrator 4	4ND71B0	2019-12-31
1 x 0.5mL	vWF Calibrator 5	5ND71B0	2019-12-31
1 x 0.5mL	vWF High Control	AND71B0	2019-12-31
1 x 0.5mL	vWF Low Control	BND71B0	2019-12-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.65	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (AND71B0)	0.84 – 1.40 IU/mL	1.16	OK
	Low Control (BND71B0)	0.18 – 0.42 IU/mL	0.31	OK

**Durchführung von  
Prepared by**

Michael KAFKA 

22. OKT. 2018  
Datum / Date:

**Freigabe von  
Approved by**

Fikret NASUFI 

22. OKT. 2018  
Datum / Date: