
	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQRK00AZ
---	---	----------

## TECHNOZYM<sup>®</sup> vWF:CBA ELISA Collagen Typ I Kit 96T

REF      5450311  
LOT      RK82B00.01  
      2019-12-31

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
1 x	vWF:CBA Plate Collagen Typ I	RNM71B1	2020-08-31
1 x 0.3mL	vWF:Ag Conjugate	RNH81B0	2020-05-31
1x 80mL	Wash buffer Concentrate	RNA82B0	2020-06-30
1 x 90mL	Incubation buffer	RNB82B0	2020-06-30
1 x 12mL	TMB Substrate	RKN82B0	2020-05-31
1 x 12mL	Stop solution	RKK81B0	2020-04-30
1 x 0.5mL	vWF:CBA Calibrator 1	1ND71B0	2019-12-31
1 x 0.5mL	vWF:CBA Calibrator 2	2ND71B0	2019-12-31
1 x 0.5mL	vWF:CBA Calibrator 3	3ND71B0	2019-12-31
1 x 0.5mL	vWF:CBA Calibrator 4	4ND71B0	2019-12-31
1 x 0.5mL	vWF:CBA Calibrator 5	5ND71B0	2019-12-31
1 x 0.5mL	vWF:CBA High Control	AND71B0	2019-12-31
1 x 0.5mL	vWF:CBA Low Control	BND71B0	2019-12-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	2.37	OK
	Calibrator 5	< 0.1	0.03	OK
Verwendbarkeit Usability	High Control (AND71B0)	0.72 – 1.20 IU/mL	0.95	OK
	Low Control (BND71B0)	0.17 – 0.41 IU/mL	0.35	OK

**Durchführung von**  
**Prepared by**

Michael KAFKA 

10. OKT. 2018

Datum / Date:

**Freigabe von**  
**Approved by**

Fikret NASUFI 

10. OKT. 2018

Datum / Date: