
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRP00AZ
---	---	----------

TECHNOZYM® ADAMTS-13 INH ELISA Kit 96T

REF 5450401
LOT RP8BB00.01
 2020-05-31

Inhalt Content	Reagenz Reagent	LOT	
1 x 12	Adamts13 INH Plate	RPS89B0	2020-07-31
1 x 0.3mL	Adamts13 INH Conjugate	RPV82B0	2020-08-31
1x 80mL	Wash buffer Concentrate	RNA82B0	2020-06-30
1 x 90mL	Incubation buffer	RNB82B0	2020-06-30
1 x 12mL	TMB Substrate	RKN82B0	2020-05-31
1 x 12mL	Stop solution	RKK82B0	2020-08-31
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT71B1	2020-08-31
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT71B1	2020-08-31
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT71B1	2020-08-31
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT71B1	2020-08-31
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT71B1	2020-08-31
1 x 0.5mL	Adamts13 INH High Control	RPY71B1	2020-08-31
1 x 0.5mL	Adamts13 INH Low Control	RPZ71B1	2020-08-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.67	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (RPY71B1)	60.24 – 100.40 U/mL	77.90	OK
	Low Control (RPZ71B1)	16.64 – 38.84 U/mL	29.79	OK

**Durchführung von
Prepared by**

Michael KAFKA 

03. OKT. 2018

Datum / Date:

**Freigabe von
Approved by**

Fikrol NASUFI 

03. OKT. 2018

Datum / Date: