

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQRP00AZ
---	---	----------

### TECHNOZYM® ADAMTS-13 INH ELISA Kit 48T

**REF**      5450451  
**LOT**      RP89B0S.02  
      2019-08-31

Inhalt Content	Reagenz Reagent	<b>LOT</b>	
1 x 6	Adamts13 INH Plate	RPS87BS	2020-06-30
1 x 0.3mL	Adamts13 INH Conjugate	RPV81B0	2020-04-30
1x 80mL	Wash buffer Concentrate	RNA82B0	2020-06-30
1 x 90mL	Incubation buffer	RNB82B0	2020-06-30
1 x 12mL	TMB Substrate	RKN82B0	2020-05-31
1 x 12mL	Stop solution	RKK81B0	2020-04-30
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH High Control	RPY71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Low Control	RPZ71B0	2019-08-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.61	OK
	Calibrator 5	< 0.1	0.03	OK
Verwendbarkeit Usability	High Control (RPY71B0)	60.24 – 100.40 U/mL	75.67	OK
	Low Control (RPZ71B0)	16.64 – 38.84 U/mL	27.18	OK

**Durchführung von  
Prepared by**

  
FIBRET NASUBI

27. SEP. 2018

Datum / Date:

**Freigabe von  
Approved by**

  
Eva Maria HAGER

27. SEP. 2018

Datum / Date: