

CERTIFICATE OF ANALYSIS

CHROMOGENIX

Instrumentation
Laboratory
A Werfen Company

PRODUCT NAME:	Antithrombin 1 x 25 IU	EXPIRATION DATE:	2020-05
PRODUCT NUMBER:	81079639		
LOT NUMBER:	N0487595		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Antithrombin Activity	25	30	IU/vial	29	PASS
Assay Slope @ 37°C	-0.0024	-0.0034		-0.0030	PASS
Intercept @ 37°C	0.40	0.60		0.51	PASS
r ²	0.990	--		0.999	PASS

PREPARED BY: Nehal Patel *N.m. Patel*
 PRINT SIGNATURE

DATE: 5/16/2018 *5-16-18*

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: Donna Tencza *D. Tencza*
 PRINT SIGNATURE

DATE: 5/16/2018 *5/16/18*

TITLE: Senior QC Analyst