

**CERTIFICATE OF ANALYSIS**

**CHROMOGENIX**


**Instrumentation  
Laboratory**  
 A Werfen Company

|                 |                   |                  |         |
|-----------------|-------------------|------------------|---------|
| PRODUCT NAME:   | FXa, 10 x 71 nkat | EXPIRATION DATE: | 2020-12 |
| PRODUCT NUMBER: | 82098539          |                  |         |
| LOT NUMBER:     | N0688432          |                  |         |

| TEST DESCRIPTION                 | SPECIFICATIONS |       |           | RESULTS | DISPOSITION |
|----------------------------------|----------------|-------|-----------|---------|-------------|
|                                  | MIN            | MAX   | UNITS     |         |             |
| FACTOR Xa, 71 nkat (650515)      |                |       |           |         |             |
| ACTIVITY                         | 61             | 81    | nkat/vial | 71      | PASS        |
| STANDARD CURVE (r <sup>2</sup> ) | 0.990          | 1.000 | --        | 1.000   | PASS        |
| CONTROL RECOVERY                 | 67.8           | 75.0  | nkat/vial | 68.3    | PASS        |

PREPARED BY: Maria Oettinger *Maria Oettinger*  
 PRINT SIGNATURE

DATE: 9/17/2018 *9-17-18*

TITLE: QC Analyst III

APPROVAL SIGNATURE: Debra Lee Harz *Debra Lee Harz*  
 PRINT SIGNATURE

DATE: 9-18-18

TITLE: QC Technical Supervisor