

# CERTIFICATE OF ANALYSIS



PRODUCT NAME :	COAMATIC FACTOR VIII	EXPIRATION DATE: 2020-07
PRODUCT NUMBER :	82258563	
LOT NUMBER:	N0789215	

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	PASS/FAIL
	MIN	MAX	UNITS		
HIGH RANGE					
Slope	0.33	0.57	--	0.41	PASS
Intercept	---	0.12	--	0.04	PASS
r	0.990	---	--	0.998	PASS
LOW RANGE					
Slope	2.8	6.9	--	4.5	PASS
Intercept	---	0.21	--	0.11	PASS
r	0.990	---	--	0.998	PASS

We hereby certify that the above product(s):  
 · have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;  
 · meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

8-30-18

PREPARED BY: Nehal Patel N.M. Patel  
 PRINT SIGNATURE

DATE: 8/30/2018

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: Debra Lee Harz Debra Lee Harz  
 PRINT SIGNATURE

DATE: 8/30/18  
8/30/2018

TITLE: Technical Supervisor