

CERTIFICATE OF ANALYSIS

CHROMOGENIX


**Instrumentation
Laboratory**
 A Werfen Company

| | | | |
|-----------------|--------------------------|------------------|---------|
| PRODUCT NAME: | COATEST APC Resistance V | EXPIRATION DATE: | 2020-01 |
| PRODUCT NUMBER: | 82312063 | | |
| LOT NUMBER: | N0880097 | | |

| TEST DESCRIPTION | SPECIFICATIONS | | UNITS | RESULTS | Pass/Fail |
|--------------------------------|----------------|-----|---------|---------|-----------|
| | MIN | MAX | | | |
| Diluted Control Plasma Level 1 | 30 | 47 | seconds | 33 | PASS |
| Diluted Control Plasma Level 2 | 31 | 51 | seconds | 34 | PASS |

| APC-V ratio on ACL | | | | | |
|------------------------|-----|-----|----|-----|------|
| Control Plasma Level 1 | 2.8 | 3.2 | -- | 3.1 | PASS |
| Control Plasma Level 2 | 1.4 | 1.9 | -- | 1.7 | PASS |

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: Nehal Patel *N.M. Patel* DATE: 9-21-18
 PRINT SIGNATURE 9/21/2018

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: Debra Lee Harz *Debra Lee Harz* DATE: 9/21/18
 PRINT SIGNATURE 9/21/2018

TITLE: Technical Supervisor