

CERTIFICATE OF ANALYSIS

CHROMOGENIX

Instrumentation
Laboratory
A Werfen Company

PRODUCT NAME:	S-2765, 25 mg Substrate	EXPIRATION DATE:	2021-08
PRODUCT NUMBER:	82141339		
LOT:	N0880084		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.3	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.2	PASS

PREPARED BY: Maria Oettinger *Maria Oettinger* DATE: 9/10/2018 *9-10-18*
 PRINT SIGNATURE
 TITLE: QC Analyst III
 APPROVAL SIGNATURE: Debra Lee Harz *Debra Lee Harz* DATE: 9-10-18
 PRINT SIGNATURE
 TITLE: QC Technical Supervisor