
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ6A0.01
---	---	----------


TECHNOCLOT® PT Owren manual 10x10 mL

REF 5005037
LOT 6M81BB0.01
 2020-02-29

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result
Verwendbarkeit Usability Ceveron® manual	CCN LOT 1P71B01	91.4 - 115.2 % 0.91 - 1.15 INR 113.6 1.02 OK
	CCA LOT 3P71B01	23.9 - 30.1 % 1.84 - 2.31 INR 25.7 2.17 OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK

Durchführung von
Prepared by

Suzana GRESAK-PODAKAR




11. JUNI 2018

Datum / Date:

Freigabe von
Approved by

Eva Maria HAGER



11. JUNI 2018

Datum / Date: