

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ9K00AZ
---	---	----------

## Ceveron<sup>®</sup> TGA BUF 5x1mL

REF      5006370  
LOT      9K82B.01  
      2020-09-30

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
5x1 mL	Ceveron <sup>®</sup> TGA BUF	9K82B	2020-09-30

Analytische Prozedur Analytical Procedure			Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls <span style="border: 1px solid black; padding: 2px;">LOT</span> 1Y82B00	Peak Thrombin nM	Ceveron <sup>®</sup> TGA RC high <span style="border: 1px solid black; padding: 2px;">LOT</span> 7Q81B	302 - 503	462.0	OK
Kontrollwiederfindung Recovery of controls <span style="border: 1px solid black; padding: 2px;">LOT</span> 2783B00	Peak Thrombin nM	Ceveron <sup>®</sup> TGA RC high <span style="border: 1px solid black; padding: 2px;">LOT</span> 7Q81B	115 - 191	174.8	OK
Etiketten und Barcode Überprüfung Label and bar code Verification				OK	

Durchführung von  
Prepared by

Fikret NASUFI *Fikret Nasufi*

15. NOV. 2018

Datum / Date:

Freigabe von  
Approved by

Michael KAFKA *[Signature]*

16. NOV. 2018

Datum / Date: