



	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ6PT00AZ
---	---	-----------

TECHNOZYM® ADAMTS-13 INH Calibrator Set 5x0.5mL

REF 5450461
LOT 6PT81B0.01
 2020-08-31

Inhalt Content	Reagenz Reagent	LOT	
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT71B1	2020-08-31
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT71B1	2020-08-31
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT71B1	2020-08-31
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT71B1	2020-08-31
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT71B1	2020-08-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.64	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (RPY71B1)	60.24 – 100.40 U/mL	79.68	OK
	Low Control (RPZ71B1)	16.64 – 38.84 U/mL	29.20	OK

**Durchführung von
Prepared by**

Michael KAFKA 

03. OKT. 2018
Datum / Date:

**Freigabe von
Approved by**

Filbert NÄSUF 

03. OKT. 2018
Datum / Date: