



| | | |
|---|---|-----------|
|  | ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS | DQ7PT00AZ |
|---|---|-----------|

TECHNOZYM[®] ADAMTS-13 INH Control Set 2x0.5mL

REF 5450463
LOT 7PT81B0.01
 2020-08-31

| Inhalt Content | Reagenz Reagent | LOT |  |
|-------------------|---------------------------|---|---|
| 1 x 0.5mL | Adamts13 INH High Control | RPY71B1 | 2020-08-31 |
| 1 x 0.5mL | Adamts13 INH Low Control | RPZ71B1 | 2020-08-31 |

| Analytische Prozedur Analytical Procedure | | Anforderung Requirement | Ergebnis Result | |
|--|------------------------|----------------------------|--------------------|----|
| OD | Calibrator 1 | > 1.0 | 1.64 | OK |
| | Calibrator 5 | < 0.1 | 0.01 | OK |
| Verwendbarkeit Usability | High Control (RPY71B1) | 60.24 – 100.40 U/mL | 79.68 | OK |
| | Low Control (RPZ71B1) | 16.64 – 38.84 U/mL | 29.20 | OK |

**Durchführung von
Prepared by**

Michael KAFKA 

03. OKT. 2018

Datum / Date:

**Freigabe von
Approved by**

Fitrot NASUPI 

03. OKT. 2018

Datum / Date: