



Certificate of Analysis

Factor IXa Control

REF	Catalog number	9588
LOT	Batch code	HKQ1542
	Use by	2023 10
	Upper limit of temperature	-20°C

Activity after reconstitution with 2.0 mL water:

Acceptance range
0.41 - 0.61 mIU/mL

The above Factor IXa activity apply only to Lot#HKQ1542.

The calibration has been made against the 2nd International Standard for FIXa, NIBSC Code 14/316 using the Rox FIX-A kit (Art No 95 00 30).

The above Lot# has been approved for release after analyses according to established protocols with results demonstrating fulfilment of product specifications.

All data concerning materials used, production data and release testing are documented and traceable under the individual Lot# which is stated on each package.



Pia Bryngelhed
Quality Manager, Rossix AB

November 06, 2018

Date



Rossix AB
Taljegårdsgatan 3B
SE-431 53 Molndal
Sweden

Phone: +46 31 706 8965
Fax: +46 31 706 8966
info@rossix.com
www.rossix.com

Rossix AB operates an ISO 13485 Medical Devices
Quality Management System certified by BSI.