

# CERTIFICATE OF ANALYSIS

**CHROMOGENIX**


**Instrumentation  
Laboratory**  
A Werfen Company

PRODUCT NAME :	COAMATIC FACTOR VIII	EXPIRATION DATE: 2020-09
PRODUCT NUMBER :	82258563	
LOT NUMBER:	N0980563	

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	PASS/FAIL
	MIN	MAX	UNITS		
<b>HIGH RANGE</b>					
Slope	0.33	0.57	--	0.42	PASS
Intercept	---	0.12	--	0.05	PASS
r	0.990	---	--	0.997	PASS
<b>LOW RANGE</b>					
Slope	2.8	6.9	--	4.1	PASS
Intercept	---	0.21	--	0.12	PASS
r	0.990	---	--	0.999	PASS

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: Nehal Patel *N. M. Patel*  
 PRINT SIGNATURE

DATE: 10-18-18  
10/18/2018

TITLE: Senior QC Analyst  
 APPROVAL SIGNATURE: Debra Lee Harz *Debra Lee Harz*  
 PRINT SIGNATURE

DATE: 10-18-18  
10/18/2018

TITLE: Technical Supervisor