

CERTIFICATE OF ANALYSIS



PRODUCT NAME:	CoaTest SP4 FVIII Kit	EXPIRATION DATE:	2020-05
PRODUCT NUMBER:	82409463		
LOT NUMBER:	N0688431		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION	
	MIN	MAX				PASS/FAIL
Coatest SP4 Factor VIII Kit (82409463)						
High Range	Slope	0.0077	0.0112	--	0.0083	PASS
	Intercept	--	0.250	--	0.130	PASS
	Standard Curve (r)	0.99	--	--	0.997	PASS
Low Range	Slope	0.020	0.060	--	0.039	PASS
	Intercept	--	0.406	--	0.165	PASS
	Standard Curve (r)	0.99	--	--	0.999	PASS

PREPARED BY: Nehal Patel *N.M. Patel* DATE: 6/27/2018 *6-27-18*
 PRINT SIGNATURE

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: Donna Tencza *D. Tencza* DATE: 6/27/2018 *6/27/18*
 PRINT SIGNATURE

TITLE: Senior QC Analyst