

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	S-2302, 25 mg Substrate	PRODUCT #:	00082034039
EXP. DATE:	2021-11	LOT #:	N0789373
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	820340

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	24.8	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.1	PASS

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
S-2302	N0789373	2021-11	3 mL

PREPARED BY: M. Oettinger *M. Oettinger* DATE: 8/10/2018 *8-10-18*

APPROVAL SIGNATURE: D. Tencza *D. Tencza* DATE: 8/10/2018 *8/10/18*