

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRXXAZ
---	---	---------

TECHNOZYM® t-PA Combi Actibind ELISA Kit 96T

REF TC16000
LOT 160002B8.01
 2020-02-29

Inhalt Content	Reagenz Reagent	LOT	
1x	anti t-PA Plate	RGN82B0	2020-08-31
1x0.3mL	POX anti t-PA	RKM81B0	2020-08-31
1x80mL	Wash buffer Concentrate	RNA82B0	2020-06-30
1x90mL	Incubation buffer	RNB82B0	2020-06-30
1x12mL	TMB Substrat	RKN82B0	2020-05-31
1x12mL	Stop solution	RKK82B0	2020-08-31
1x0.5mL	Standard t-PA	RAW11B4	2020-02-29
1x20mL	Detection mix buffer- TRIS	RDA81B0	2020-09-30
1x	Detection mixture	RBA81B0	2020-10-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1 activity	> 0.7	0.92	OK
	Calibrator 5 activity	< 0.1	0.04	OK
	Calibrator 1 antigen	> 1.0	1.35	OK
	Calibrator 5 antigen	< 0.1	0.00	OK

**Durchführung von
Prepared by**

Fikrot NASUFI *Fikrot Nasufi*

11. OKT. 2018
Datum / Date:

**Freigabe von
Approved by**

Michael KAFKA *[Signature]*

11. OKT. 2018
Datum / Date: