
	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ7Q00AZ
---	---	----------

## Ceveron<sup>®</sup> TGA RC High 5x1 mL

REF      5006227  
LOT      7Q81B.01  
      2019-07-31

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
5x1 mL	Ceveron <sup>®</sup> TGA RC High	7Q81B	31-Jul-2019

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	Ceveron <sup>®</sup> TGA Control High (nM Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 1Y82B00	302 – 503	479.2	OK
	Ceveron <sup>®</sup> TGA Control Low (nM Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 2783B00	115 – 191	184.1	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

Michael KAFKA 

01. OKT. 2018

Datum / Date:

Freigabe von  
Approved by

Fikret NASUFI 

01. OKT. 2018

Datum / Date: