

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0W00.01
---	---	-----------


Technochrom AT III Kit ~ 100 T.

REF	5340224
LOT	0W82B00.01
	2020-12-31

Inhalt Content	Reagenz Reagent	LOT	
1x	AT III Substrate Th-1	8B81B00	2021-03-31
1x	Technochrom ATIII Reagent A2	7S71B00	2020-12-31
2x25mL	Natrium Chlorid Lösung 0.9%	9581B00	2023-02-28

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha Chromogenic Substrate F IIa	CCN LOT 1P71B00	75.8 – 126.3 %	102.5	OK
	CCA LOT 3P71B00	33.3 – 55.4 %	43.9	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Suzana GRESAK PODAKAR

11. JUNI 2018

Datum / Date:

**Freigabe von
Approved by**


Christoph ENNE

11. JUNI 2018

Datum / Date: