

CERTIFICATE OF ANALYSIS




Instrumentation  
Laboratory  
A Werfen Company

|                 |                       |                  |         |
|-----------------|-----------------------|------------------|---------|
| PRODUCT NAME:   | CoaTest SP4 FVIII Kit | EXPIRATION DATE: | 2020-07 |
| PRODUCT NUMBER: | 82409463              |                  |         |
| LOT NUMBER:     | N0789377              |                  |         |

| TEST DESCRIPTION                              | SPECIFICATIONS |        |       | RESULTS | DISPOSITION |
|---|----------------|--------|-------|---------|-------------|
|   | MIN            | MAX    | UNITS |         |             |
| <b>Coatest SP4 Factor VIII Kit (82409463)</b> |                |        |       |         |             |
| <b>High Range</b>                             |                |        |       |         |             |
| Slope   | 0.0077         | 0.0112 | --    | 0.0077  | PASS        |
| Intercept                                     | --             | 0.250  | --    | 0.110   | PASS        |
| Standard Curve (r)                            | 0.99           | --     | --    | 0.998   | PASS        |
| <b>Low Range</b>                              |                |        |       |         |             |
| Slope   | 0.020          | 0.060  | --    | 0.038   | PASS        |
| Intercept                                     | --             | 0.406  | --    | 0.143   | PASS        |
| Standard Curve (r)                            | 0.99           | --     | --    | 0.999   | PASS        |

PREPARED BY: D.Tencza  DATE: 8/21/2018 8/21/18  
 PRINT (SIGNATURE)

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: D.Harz  DATE: 8/21/2018 8/21/18  
 PRINT (SIGNATURE)

TITLE: Hemostasis Technical Supervisor