

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1000.01
---	---	-----------

TECHNOVIEW Apixaban CAL 5x1mL

REF 5090269
LOT 1082B00.01
 2020-09-30

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	TECHNOVIEW Apixaban Cal 1	1181B00	2020-09-30
1x1mL	TECHNOVIEW Apixaban Cal 2	1281B00	2020-09-30
1x1mL	TECHNOVIEW Apixaban Cal 3	1381B00	2020-09-30
1x1mL	TECHNOVIEW Apixaban Cal 4	1481B00	2020-09-30
1x1mL	TECHNOVIEW Apixaban Cal 5	1581B00	2020-09-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	TECHNOVIEW Apixaban Cont Low LOT 3971B00	74.9 – 139.1 ng/mL	100.4	OK
	TECHNOVIEW Apixaban Cont High LOT 3D71B00	259.5 – 432.5 ng/mL	302.7	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Suzana GRESAK PODAKAR


03. DEZ. 2018

Datum / Date:

Freigabe von
Approved by


Stefan STEINER

03. DEZ. 2018

Datum / Date: