

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ9600.01
---	---	-----------

### Ceveron® Cleaning Solution 3% 25 mL

REF            9820300


LOT            9682B00.01

            2019-08-31

Analytische Verfahren Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	11.5 – 12.5	11.7	OK
Volumen Volume	25.0 – 26.5 mL	26.0	OK
Sterilität Sterility	0 CFU/mL	0	OK


Verwendbarkeit, Fibrinogen, Clauss method Ceveron® alpha Usability, Fibrinogen, Clauss method Ceveron® alpha	CCN LOT 1P71B01	3.17 – 4.75 g/L	4.27	OK
	CCA LOT 3P71B01	1.02 – 1.54 g/L	1.31	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

  
**Eva Maria HAGER**

20. AUG. 2018  
Datum / Date:

Freigabe von  
Approved by

  
**Stefan STEINER**

21. AUG. 2018  
Datum / Date: