

CERTIFICATE OF ANALYSIS


CHROMOGENIX


**Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME:	<u>COATEST APC Resistance Kit</u>	EXPIRATION DATE:	<u>2020-03</u>
PRODUCT NUMBER:	<u>82264363</u>		
LOT NUMBER :	<u>N1182205</u>		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	Pass/Fail
	MIN	MAX			
APC-ratio on ACL					
Control Plasma Level 1 (Lot: _____ N0864887)	3.5	4.4	--	4.0	PASS
Control Plasma Level 2 (Lot: _____ N0386562)	--	2.3	--	2.1	PASS

We hereby certify that the above product(s):
 - have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
 - meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: D.Tencza 
 PRINT SIGNATURE

DATE: 12/11/2018 ^{12/11/18}

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: D.Harz 
 PRINT SIGNATURE

DATE: 12/11/2018 ^{12/11/18}

TITLE: Technical Supervisor