

CERTIFICATE OF ANALYSIS

CHROMOGENIX



**Instrumentation
Laboratory**

A Werfen Company

PRODUCT NAME:	CoaTest SP Factor FVIII Kit	EXPIRATION DATE:	2020-07
PRODUCT NUMBER:	82408663		
LOT NUMBER :	N0789376		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
High Range					
Slope	0.0077	0.0112	--	0.0088	PASS
Intercept	--	0.250	--	0.110	PASS
Standard Curve (r)	0.99	--	--	0.998	PASS
Low Range					
Slope	0.020	0.060	--	0.038	PASS
Intercept	--	0.406	--	0.145	PASS
Standard Curve (r)	0.99	--	--	0.998	PASS

PREPARED BY: Maria Oettinger *Maria Oettinger* DATE: 08/21/18 *8-21-18*
 PRINT SIGNATURE

TITLE QC Analyst III

APPROVAL SIGNATURE: Debra Lee Hays *Debra Lee Hays* DATE: 08/21/18 *8/21/18*
 PRINT SIGNATURE