

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT: S-2302, 25 mg Substrate	PRODUCT #: 00082034039
EXP. DATE: 2021-08	LOT #: N0487244
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 820340

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	24.9	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.0	PASS

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
S-2302	N0487244	2021-08	3 mL

PREPARED BY:

Maria Dittler

DATE:

5-3-18

APPROVAL SIGNATURE:

J. Amca

DATE:

5/3/18