

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ8F00AZ
---	---	----------

TECHNOTHROMBIN® TGA SUB 50x1.5mL

REF 5006230

LOT 8F71B01.02

 2020-08-31

Inhalt Content	Reagenz Reagent	LOT	
50x1.5mL	TGA Substrat	8F71B01	2020-08-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z82B00	TGA RCH Reagent LOT 7F81B00	377 - 628	492.9 OK
	TGA Control Low (nM Peak Thrombin) LOT 2883B00	TGA RCH Reagent LOT 7F81B00	206 - 344	259.0 OK

Durchführung von
Prepared by

Michael KAFKA 

17. DEZ. 2018

Datum / Date:

Freigabe von
Approved by

Fikret NASUFI 

18. DEZ. 2018

Datum / Date: