
	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ3D00.01
---	---	-----------

**TECHNOVIEW® Apixaban CONT High 5x1mL**

**REF**            5090270  
**LOT**            3D81B00.01  
            2020-10-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		229.3 – 382.1 ng/mL	305.7	OK
Verwendbarkeit Usability Ceveron ® alpha	Anti-Xa Apixaban	229.3 – 382.1 ng/mL	280.1	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von  
Prepared by**

  
**Suzana GRESAK-PODAKAR**

25. JAN. 2019  
Datum / Date:

**Freigabe von  
Approved by**

  
**Stefan STEINER**

25. JAN. 2019  
Datum / Date: