
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3900.01
---	---	-----------

TECHNOVIEW® Apixaban CONT Low 5x1ml

REF 5090271
LOT 3981B00.01
 2020-11-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		64.2 – 119.2 ng/mL	91.7	OK
Verwendbarkeit Usability Ceveron® alpha	Anti-Xa Apixaban	64.2 – 119.2 ng/mL	79.6	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Suzana GRESAK PODAKAR

25. JAN. 2019
Datum / Date:

**Freigabe von
Approved by**


Stefan STEINER

25. JAN. 2019
Datum / Date: