



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ0F00.03

F VIII Inhibitor Reagent Kit 2-4T.

REF 5152005
LOT 0F91B00.01
⌚ 2020-06-30

Inhalt Content	Reagenz Reagent	LOT	⌚
2 x 2.63 mL	FVIII Normal Plasma	1G81B00	2020-06-30
1 x 1 mL	FVIII Inhibitor Plasma	2K82B00	2020-12-31
1 x 1 mL	FVIII Inhibitor Free Plasma	2M82B0D	2020-12-31
1 x 17 mL	Imidazol Puffer	9J84BA0	2023-07-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability FVIII clotting, Siron LS Ceveron® alpha	FVIII Inhibitor Plasma LOT 2K82B00	19.0 – 31.6 BE/mL	23.7	OK
	FVIII Inhibitor Free Plasma LOT 2M82B0D	< 0.4 BE/mL	< 0.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Suzana GRESAK PODAKAR

13. FEB. 2019

Datum / Date:

Freigabe von
Approved by

Stefan STEINER

13. FEB. 2019

Datum / Date: