

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQRP00AZ
---	---	----------

**TECHNOZYM<sup>®</sup> ADAMTS-13 INH ELISA Kit 96T**

REF      5450401  
LOT      RP93B00.01  
      2020-08-31

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
1 x 12	Adamts13 INH Plate	RPS91B0	2021-01-31
1 x 0.3mL	Adamts13 INH Conjugate	RPV82B0	2020-08-31
1x 80mL	Wash buffer Concentrate	RNA83B0	2020-10-31
1 x 90mL	Incubation buffer	RNB83B0	2020-10-31
1 x 12mL	TMB Substrate	RKN83B0	2020-12-31
1 x 12mL	Stop solution	RKK82B0	2020-08-31
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT91B0	2021-01-31
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT91B0	2021-01-31
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT91B0	2021-02-28
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT91B0	2021-02-28
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT91B0	2021-01-31
1 x 0.5mL	Adamts13 INH High Control	RPY91B0	2021-02-28
1 x 0.5mL	Adamts13 INH Low Control	RPZ91B0	2021-02-28

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.66	OK
	Calibrator 5	< 0.1	0.00	OK
Verwendbarkeit Usability	High Control (RPY91B0)	56.55 – 94.25 U/mL	79.43	OK
	Low Control (RPZ91B0)	15.64 – 36.48 U/mL	28.30	OK

**Durchführung von  
Prepared by**

Michael KAFKA 

18. FEB. 2019  
Datum / Date:

**Freigabe von  
Approved by**

Fikret NASUFI 

18. FEB. 2019  
Datum / Date: