



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRP00AZ

TECHNOZYM® ADAMTS-13 INH ELISA Kit 48T

REF 5450451
LOT RP94B0S.01
2020-08-31

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x 6	Adamts13 INH Plate	RPS8BBS	2020-12-31
1 x 0.3mL	Adamts13 INH Conjugate	RPV82B0	2020-08-31
1x 80mL	Wash buffer Concentrate	RNA83B0	2020-10-31
1 x 90mL	Incubation buffer	RNB83B0	2020-10-31
1 x 12mL	TMB Substrate	RKN83B0	2020-12-31
1 x 12mL	Stop solution	RKK82B0	2020-08-31
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT91B0	2021-01-31
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT91B0	2021-01-31
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT91B0	2021-02-28
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT91B0	2021-02-28
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT91B0	2021-01-31
1 x 0.5mL	Adamts13 INH High Control	RPY91B0	2021-02-28
1 x 0.5mL	Adamts13 INH Low Control	RPZ91B0	2021-02-28

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.60	OK
	Calibrator 5	< 0.1	0.04	OK
Verwendbarkeit Usability	High Control (RPY91B0)	56.55 – 94.25 U/mL	71.29	OK
	Low Control (RPZ91B0)	15.64 – 36.48 U/mL	26.75	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

20. FEB. 2019
Datum / Date:

Freigabe von
Approved by

Michael KAFKA *Michael Kafka*

20. FEB. 2019
Datum / Date: