

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7PT00AZ
---	---	-----------

TECHNOZYM® ADAMTS-13 INH Control Set 2x0.5mL

REF 5450463
LOT 7PT81B0.02
 2020-08-31

Inhalt Content	Reagenz Reagent	LOT	
1 x 0.5mL	Adamts13 INH High Control	RPY71B1	2020-08-31
1 x 0.5mL	Adamts13 INH Low Control	RPZ71B1	2020-08-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	High Control (RPY71B1)	60.24 – 100.40 U/mL	79.41	OK
	Low Control (RPZ71B1)	16.64 – 38.84 U/mL	29.58	OK

**Durchführung von
Prepared by**

Michael KAFKA 

15. NOV. 2018
Datum / Date:

**Freigabe von
Approved by**

Fikret NASUFI 

15. NOV. 2018
Datum / Date: